



MULTIPLIER FACTORS OF QUENCHANT SOLUTIONS FOR COMMONLY USED REFRACTOMETERS

PRODUCT	IFT 40	Reichert 10440	Atago Decimal	Atago Digital (Pal 1)
UCON A	2.0	2.0	1.78	1.78
UCON HT	2.0	2.0	1.78	1.78
UCON C	2.0	2.0	1.78	1.78
UCON B	2.7	2.7	2.5	2.5
UCON E	2.7	2.7	2.5	2.5
UCON RL	2.7	2.7	2.5	2.5
UCON NTNN	7.2	7.2	7.0	7.0
TQU 1500	3.35	3.35	3.33	3.33
TQU 2000	10.0	10.0	9.1	9.1
TQU 2500	8.0	8.0	8.0	8.0